

Human IL-2 Sandwich ELISA Kit Datasheet

Please read it entirely before use

Catalogue Number: KE00017

Size: 5*96T

Sensitivity: 3.6 pg/mL

Range: 15.6-1000 pg/mL

Usage: For the quantitative detection of human IL-2 concentrations in serum, plasma and cell culture supernatant.

This product is for research use only and not for use in human or animal therapeutic or diagnostic.

Table of content	page
1. Background	3
2. Principle	3
3. Required Materials	3
4. Kit Components and Storage	4
5. Safety Notes	4
6. Sample Collection and Storage	4
7. Regent Preparation	5
8. Assay Procedure Summary	6
9. Validation Data	7
9.1 Standard curve	7
9.2 Precision	8
9.3 Recovery	8
9.4 Sample values	8
9.5 Sensitivity	8
9.6 Linearity	9
9.7 Calibration	9
10. References	9

1. Background

Interleukin-2 (IL-2) is a four-helix bundle, type I cytokine that functions as a growth factor for a wide range of leukocytes. In the immune system, IL-2 is essential for immune homeostasis, normal T regulatory cell function, and self-tolerance. It regulates immune cell homeostasis and has been used to treat a range of disorders including cancer and autoimmune disease. IL-2 signals through heterodimerization of the IL-2R β and IL-2R γ receptor subunits.

2. Principle



Sandwich ELISA structure (Detection antibody labeled with biotin)

A capture antibody is pre-coated onto the bottom of wells which binds to analyte of interest. A detection antibody labeled with biotin also binds to the analyte. Streptavidin-HRP binds to the biotin. TMB acts as the HRP substrate and the solution color will change from colorless to blue. A stop solution containing sulfuric acid turns solution yellow. The color intensity is proportional to the quantity of bound protein which is measurable at 450 nm with the correction wavelength set at 630 nm.

3. Required Materials

- 3.1 A microplate reader capable of measuring absorbance at 450 nm with the correction wavelength set at 630 nm.
- 3.2 Calibrated, adjustable precision pipettes and disposable plastic tips. A manifold multi-channel pipette is recommended for large assays.
- 3.3 Plate washer: automated or manual.
- 3.4 Absorbent paper towels.
- 3.5 Glass or plastic tubes to prepare standard and sample dilutions.
- 3.6 Beakers and graduated cylinders.
- 3.7 Log-log or semi-log graph paper or computer and software for ELISA data analysis. A four-parameter logistic (4-PL) curve-fit is recommended.

4. Kit Components and Storage

Microplate - antibody coated 96-well microplate (8 well × 12 strips)	5 plates	Unopened Kit: Store at 2-8°C for 6 months or -20°C for 12 months. Opened Kit: All reagents stored at 2-8°C for 7 days. Please use a new standard for each assay.
Protein standard - 2000 pg/bottle; lyophilized	10 bottles	
Detection antibody, biotinylated(100×) - 600 μ L/vial*	1 vial	
Streptavidin-horseradish peroxidase (HRP) (100×) - 600 μ L/vial*	1 vial	
Sample Diluent PT 1-ef - 150 mL/bottle. For serum and plasma	1 bottle	
Sample Diluent PT 1-df - 150 mL/bottle. For cell culture supernatant	1 bottle	
Detection Diluent - 150 mL/bottle	1 bottle	
Wash Buffer Concentrate (20×) - 150 mL/bottle	1 bottle	
Tetramethylbenzidine Substrate (TMB) - 60 mL/bottle	1 bottle	
Stop Solution - 60 mL/bottle	1 bottle	
Plate Cover Seals	15 pieces	

* Centrifugation immediately before use

5. Safety Notes

- 5.1 Avoid any skin and eye contact with Stop Solution and TMB. In case of contact, wash thoroughly with water.
- 5.2 Do not use the kit after the expiration date.
- 5.3 Do not mix or substitute reagents or materials from other kit lots or other sources.
- 5.4 Be sure to wear protective equipment such as gloves, masks and goggles during the experiment.
- 5.5 When using an automated plate washer, adding a 30 second soak period following the addition of Wash Buffer to improve assay precision

6. Sample Collection and Storage

- 6.1 Serum: Allow blood samples to clot for 30 minutes, followed by centrifugation for 15 minutes at 1000xg. Clear serum can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.
- 6.2 Plasma: Use EDTA, heparin, or citrate as an anticoagulant for plasma collection. Centrifuge for 15 minutes at 1000xg within 30 minutes of collection. The plasma can be assayed immediately or aliquoted and stored at -20°C. Avoid repeated freeze-thaw cycles.
- 6.3 Cell Culture Supernatant: Remove particulates by centrifugation for 5 minutes at 500xg and assay immediately or aliquot and store samples at ≤ -20°C. Avoid repeated freeze-thaw cycles.

7. Regent Preparation

7.1 Wash Buffer (1X): If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Add 30 mL of Wash Buffer Concentrate(20X) to 570 mL deionized or distilled water to prepare 1X Wash Buffer.

7.2 Detection Antibody (1X): Dilute 100X Detection Antibody 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 μ L 100X Detection Antibody + 990 μ L Detection Diluent (Centrifuge the 100 X Detection Antibody solution for a few seconds prior to use).


7.3 Streptavidin-HRP (1X): Dilute 100X Streptavidin-HRP 1:100 using Detection Diluent prior to assay. Suggested 1:100 dilution: 10 μ L 100X Streptavidin-HRP + 990 μ L Detection Diluent (Centrifuge the 100X Streptavidin-HRP solution for a few seconds prior to use).

7.4 Sample Dilution: Different samples should be diluted with corresponding Sample Diluent, samples may require further dilution if the readout values are higher than the highest standard OD reading. Variations in sample collection, processing and storage may affect the results of the measurement.

Recommended Dilution for different sample types: 1:2 or 1:4 is recommended for serum and plasma.

7.5 Standard Serial Dilution:

For human serum and plasma samples, add 2 mL Sample Diluent PT 1-ef in protein standard; For cell culture supernatant, add 2 mL Sample Diluent PT 1-df in protein standard.



Add # μ L of Standard diluted in the previous step	—	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L
# μ L of Sample Diluent PT 1-ef or PT 1-df	2000 μL	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L	500 μ L
	"sd7"	"sd6"	"sd5"	"sd4"	"sd3"	"sd2"	"sd1"

8. Assay Procedure Summary

Bring all reagents to room temperature before use (Detection antibody and Streptavidin-HRP can be used immediately). To avoid cross-contamination, change pipette tips between additions of each standard level, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.

8.1 Take out the required number of microplate strips and return excess strips to the foil pouch containing the drying reagent pack and reseal; store at 4°C immediately. Microplate strips should be used in one week.

8.2 Preset the layout of the microplate, including control group, standard group and sample group, add 100 μ L of each standard and sample to the appropriate wells. (Make sure sample addition is uninterrupted and completed within 5 to 10 minutes, it is recommended to assay all standards, controls, and samples in duplicate).

8.3 Seal plate with cover seal, pressing it firmly onto top of microwells. Incubate the plate for 2 hours at room temperature.

8.4 Wash

1) Gently remove the cover seal. Discard the liquid from wells by aspirating or decanting. Remove any residual solution by tapping the plate a few times on fresh paper towels.

2) Wash 4 times with 1X Wash Buffer, using at least 350-400 μ L per well. Following the last wash, firmly tap plates on fresh towels 10 times to remove residual Wash Buffer. Avoid getting any towel fibers in the wells or wells drying out completely.

8.5 Add 100 μ L of 1X Detection Antibody solution (refer to Reagent Preparation 7.2) to each well. Seal plate with cover seal and incubate for 1 hour at room temperature.

8.6 Repeat wash step in 8.4.

8.7 Add 100 μ L of 1X Streptavidin-HRP solution (refer to Reagent Preparation 7.3) to each well. Seal plate with cover seal and incubate the plate for 40 minutes at room temperature.

8.8 Repeat wash step in 8.4.

8.9 Signal development: Add 100 μ L of TMB substrate solution to each well, protected from light. Incubate for 15 to 20 minutes. Substrate Solution should remain colorless until added to the plate.

8.10 Quenching color development: Add 100 μ L of Stop Solution to each well in the same order as addition of the TMB substrate. Mix by tapping the side of the plate gently. NB: Avoid skin and eye contact with the Stop solution.

8.11 Read results: Immediately after adding Stop solution read the absorbance on a microplate reader at a wavelength of 450 nm. If possible, perform a double wavelength readout (450 nm and 630 nm).


8.12 Data analysis: Calculate the average of the duplicate readings (OD value) for each standard and sample, and subtract the average of the zero standard absorbance. Construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis, use four-parameter logistic curve-fit (4-PL) analysis to do this. If the samples have been diluted, the OD readout from the standard curve must be multiplied by the dilution factor used.

Step	Reagent	Volume	Incubation	Wash	Notes
1	Standard and Samples	100 μ L	120 min	4 times	Cover Wells incubate at room temperature
2	Diluent Antibody Solution	100 μ L	60 min	4 times	Cover Wells incubate at room temperature
3	Diluent HRP Solution	100 μ L	40 min	4 times	Cover Wells incubate at room temperature
4	TMB Substrate	100 μ L	15-20 min	Do not wash	Incubate in the dark at room temperature
5	Stop Solution	100 μ L	0 min	Do not wash	-
6	Read plate at 450 nm and 630 nm immediately after adding Stop solution. DO NOT exceed 5 minutes.				


9. Validation Data

9.1 Standard curve

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.



(pg/mL)	O.D	Average	Corrected
0	0.033 0.033	0.033	-
15.63	0.093 0.093	0.093	0.06
31.25	0.151 0.153	0.152	0.119
62.5	0.288 0.266	0.277	0.244
125	0.511 0.519	0.515	0.482
250	0.84 0.835	0.8375	0.8045
500	1.356 1.407	1.3815	1.3485
1000	2.096 2.185	2.1405	2.1075



(pg/mL)	O.D	Average	Corrected
0	0.03 0.029	0.0295	-
15.6	0.143 0.101	0.122	0.0925
31.25	0.19 0.149	0.1695	0.14
62.5	0.344 0.246	0.295	0.2655
125	0.604 0.478	0.541	0.5115
250	0.956 0.725	0.8405	0.811
500	1.467 1.319	1.393	1.3635
1000	2.026 1.997	2.0115	1.982

9.2 Precision

Intra-assay Precision (Precision within an assay) Three samples of known concentration were tested 20 times on one plate to assess intra-assay precision.

Inter-assay Precision (Precision between assays) Three samples of known concentration were tested in 24 separate assays to assess inter-assay precision.

Intra-assay Precision					Inter-assay Precision				
Sample	n	Mean (pg/mL)	SD	CV%	Sample	n	Mean (pg/mL)	SD	CV%
1	20	1109.7	35.8	3.2	1	24	1370.3	126.4	9.2
2	20	339.3	15.7	4.6	2	24	420.6	32.5	7.7
3	20	70.8	2.0	2.9	3	24	79.9	6.5	8.1

9.3 Recovery

The recovery of human IL-2 spiked to three different levels throughout the range of the assay in various matrices was evaluated.

Sample Type		Average% of Expected	Range (%)
Human plasma	1:2	88	81-96
	1:4	95	84-107
Cell culture supernatant	1:2	118	116-121
	1:4	107	103-109

9.4 Sample values

Twenty-four serum and plasma samples from volunteers were evaluated for human IL-2 in this assay. All samples measured less than the lowest standard, 15.63 pg/mL. No medical histories were available for the donors used in this study.

9.5 Sensitivity

The minimum detectable dose of human IL-2 is 3.6 pg/mL. This was determined by adding two standard deviations to the concentration corresponding to the mean O.D. of 20 zero standard replicates.

9.6 Linearity

To assess the linearity of the assay, three samples were spiked with high concentrations of human IL-2 in various matrices and diluted with the appropriate Sample Diluent to produce samples with values within the dynamic range of the assay. (The plasma samples were initially diluted 1:1)

		Human plasma (Sample Diluent PT1-ef)	Cell culture supernatant (Sample Diluent PT1-df)
1:2	Average% of Expected	83	120
	Range (%)	80-86	115-124
1:4	Average% of Expected	97	117
	Range (%)	93-101	111-121
1:8	Average% of Expected	106	110
	Range (%)	97-114	99-122
1:16	Average% of Expected	102	108
	Range (%)	98-106	102-114

9.7 Calibration

This immunoassay is calibrated against a highly purified E. coli-expressed recombinant human IL-2 produced at Proteintech Systems.

NIBSC/WHO International standard for IL-2(86/500), which was intended as a potency standard, was evaluated in this kit. The dose response curve of the International Standard(86/500) parallels the Proteintech standard curve. To convert sample values obtained with the Human IL-2 ELISA kit to approximate NIBSC (86/500) units, use the equation below.

NIBSC (86/500) approximate value (IU/mL) = 0.0170 x Proteintech Human IL-2 value (pg/mL)

10. References

1. Spangler JB. et al. (2015). Immunity. 42: 815-25.
2. Rosenberg SA. et al. (2014). J Immunol. 192: 5451-8.
3. Nelson BH. et al. (2004). J Immunol. 172: 3983-8.
4. Turka LA. et al. (2008). Front Biosci. 13: 1440-6.